



Idaho State Police

Serving since 1939

Forensic Services



Colonel G. Jerry Russell
Director

C.L. "Butch" Otter
Governor

The following are the ISP uncertainty of measurement calculations for the breath testing instruments used within the state of Idaho.

Instrument	0.04	0.08	0.200
ALCOSENSORIII*	+/- 3.0%	+/- 3.0%	+/- 3.0%
INTOX5000/EN*	+/- 3.0%	+/- 3.0%	+/- 3.0%
LIFELOC FC20	+/- 3.16%	+/- 1.81%	+/- 2.64%

*It should be noted, that ISP recognizes a +/- 3% uncertainty of measurement for the alcoseensor series and the Intoxilyzer 5000 series of instruments, at all levels, based upon specifications provided by the manufacturer through training. ISP's own data indicates a lower uncertainty, but defers to the manufacturers study for these stated values.

Jeremy Johnston
ISP Forensic Scientist III
Alcohol Discipline Leader
Idaho State Police Forensic Services